

rteres

Instrument :
FID_G
ClientSampleId :
MDL-WATER-03-QT2-2023

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/10/2023
Supervised By :Ankita Jodhani 04/10/2023

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG04072
Data File : FG011742.D
Signal(s) : FID1A.ch
Acq On : 07 Apr 2023 18:02
Sample : 02213-09
Misc :
ALS Vial : 18 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031523.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	4.571	4.544	4.607	BB	13379	139457	7.06%	4.181%
2	6.742	6.713	6.775	BB	13664	138103	6.99%	4.140%
3	8.574	8.543	8.613	BB	9355	106837	5.41%	3.203%
4	10.186	10.157	10.240	BB	10541	135969	6.88%	4.076%
5	11.633	11.597	11.673	BB	11751	145647	7.37%	4.367%
6	12.946	12.908	12.983	BB	11749	143059	7.24%	4.289%
7	14.147	14.104	14.188	BB	11000	135966	6.88%	4.076%
8	15.050	15.005	15.115	BB	141471	1975569	100.00%	59.229%
9	15.254	15.208	15.294	BB	10979	147985	7.49%	4.437%
10	16.278	16.235	16.322	BB	10239	134061	6.79%	4.019%
11	17.230	17.185	17.263	BB	9321	132828	6.72%	3.982%
Sum of corrected areas:						3335479		

FG031523.M Mon Apr 10 04:10:50 2023