

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG041019\
 Data File : FG004637.D
 Signal(s) : FID1A.ch
 Acq On : 10 Apr 2019 09:42
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 11 01:28:41 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG032719.M
 Quant Title :
 QLast Update : Wed Mar 27 13:30:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.412	4737867	45.933 ug/ml
Target Compounds			
2) N-DECANE	4.144	5968418	56.543 ug/ml
3) N-DODECANE	6.167	5951026	56.484 ug/ml
4) N-TETRADECANE	7.960	5883463	54.302 ug/ml
5) N-HEXADECANE	9.561	5806297	51.897 ug/ml
6) N-OCTADECANE	11.001	5685214	50.123 ug/ml
7) N-EICOSANE	12.312	5528306	48.360 ug/ml
8) N-DOCOSANE	13.511	5427672	47.481 ug/ml
10) N-TETRACOSANE	14.613	5390221	46.662 ug/ml
11) N-HEXACOSANE	15.636	5347552	46.649 ug/ml
12) N-OCTACOSANE	16.587	5378692	46.493 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG041019\
Data File : FG004637.D
Signal(s) : FID1A.ch
Acq On : 10 Apr 2019 09:42
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 11 01:28:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG032719.M
Quant Title :
QLast Update : Wed Mar 27 13:30:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

