

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG041321\
 Data File : FG007817.D
 Signal(s) : FID1A.ch
 Acq On : 13 Apr 2021 11:16
 Operator : DD\AJ
 Sample : PB135689BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 PB135689BS

Manual Integrations
 APPROVED

Ankita
 4/14/2021 11:06:23 AM

Integration File: autoint1.e
 Quant Time: Apr 14 01:56:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040521.M
 Quant Title :
 QLast Update : Tue Apr 06 03:05:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.770	3014985	20.189 ug/ml
Target Compounds			
1) N-OCTANE	2.372	2929427	18.303 ug/ml
2) N-DECANE	4.481	3449153	20.803 ug/ml
3) N-DODECANE	6.508	3563587	21.238 ug/ml
4) N-TETRADECANE	8.308	3641554	21.160 ug/ml
5) N-HEXADECANE	9.913	3649229	21.234 ug/ml
6) N-OCTADECANE	11.356	3677411	21.063 ug/ml
7) N-EICOSANE	12.670	3700022	21.345 ug/ml
8) N-DOCOSANE	13.870	3626678	21.070 ug/ml
10) N-TETRACOSANE	14.975	3499250	20.473 ug/ml
11) N-HEXACOSANE	15.996	3357807	20.175 ug/ml
12) N-OCTACOSANE	16.946	3225554	19.832 ug/ml
13) N-TRIACONTANE	17.833	3090139	19.666 ug/ml
14) N-DOTRIACONTANE	18.667	2988761	19.579 ug/ml
15) N-TETRATRIACONTANE	19.451	2824463	21.247 ug/ml
16) N-HEXATRIACONTANE	20.191	2757615	23.693 ug/ml
17) N-OCTATRIACONTANE	20.950	2774498	26.490 ug/ml
18) N-TETRACONTANE	21.904	2736447	27.603 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG041321\
Data File : FG007817.D
Signal(s) : FID1A.ch
Acq On : 13 Apr 2021 11:16
Operator : DD\AJ
Sample : PB135689BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
FID_G
ClientSampled :
PB135689BS

Manual Integrations
APPROVED

Ankita
4/14/2021 11:06:23 AM

Integration File: autoint1.e
Quant Time: Apr 14 01:56:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040521.M
Quant Title :
QLast Update : Tue Apr 06 03:05:53 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

