

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG041423\
 Data File : FG011798.D
 Signal(s) : FID1A.ch
 Acq On : 13 Apr 2023 09:41
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Apr 14 03:42:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG041323.M
 Quant Title :
 QLast Update : Wed Apr 12 18:20:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.040	5493480	53.765 ug/ml
Target Compounds			
2) N-DECANE	4.565	7142760	57.207 ug/ml
3) N-DODECANE	6.737	7245695	58.176 ug/ml
4) N-TETRADECANE	8.568	7178817	59.311 ug/ml
5) N-HEXADECANE	10.179	7172113	57.949 ug/ml
6) N-OCTADECANE	11.625	6986040	56.960 ug/ml
7) N-EICOSANE	12.939	6724201	55.738 ug/ml
8) N-DOCOSANE	14.139	6311066	54.557 ug/ml
10) N-TETRACOSANE	15.245	6237613	53.698 ug/ml
11) N-HEXACOSANE	16.268	6017053	53.470 ug/ml
12) N-OCTACOSANE	17.221	5887081	53.663 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG041423\
Data File : FG011798.D
Signal(s) : FID1A.ch
Acq On : 13 Apr 2023 09:41
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Apr 14 03:42:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG041323.M
Quant Title :
QLast Update : Wed Apr 12 18:20:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

