

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG041621\
 Data File : FG007857.D
 Signal(s) : FID1A.ch
 Acq On : 16 Apr 2021 09:15
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 16 21:06:36 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040521.M
 Quant Title :
 QLast Update : Tue Apr 06 03:05:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.784	7072477	47.360 ug/ml
Target Compounds				
2)	N-DECANE	4.484	8657749	52.219 ug/ml
3)	N-DODECANE	6.514	8752824	52.164 ug/ml
4)	N-TETRADECANE	8.315	8897818	51.704 ug/ml
5)	N-HEXADECANE	9.921	8772945	51.048 ug/ml
6)	N-OCTADECANE	11.368	8743688	50.081 ug/ml
7)	N-EICOSANE	12.681	8506863	49.075 ug/ml
8)	N-DOCOSANE	13.883	8231680	47.825 ug/ml
10)	N-TETRACOSANE	14.989	7955785	46.547 ug/ml
11)	N-HEXACOSANE	16.012	7598081	45.652 ug/ml
12)	N-OCTACOSANE	16.963	7366905	45.294 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG041621\
Data File : FG007857.D
Signal(s) : FID1A.ch
Acq On : 16 Apr 2021 09:15
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 16 21:06:36 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040521.M
Quant Title :
QLast Update : Tue Apr 06 03:05:53 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

