

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG042219\
 Data File : FG004764.D
 Signal(s) : FID1A.ch
 Acq On : 22 Apr 2019 11:29
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 22 22:16:06 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG032719.M
 Quant Title :
 QLast Update : Wed Mar 27 13:30:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.408	5126371	49.700 ug/ml
Target Compounds			
2) N-DECANE	4.132	6925967	65.615 ug/ml
3) N-DODECANE	6.156	6860590	65.117 ug/ml
4) N-TETRADECANE	7.950	6704102	61.876 ug/ml
5) N-HEXADECANE	9.552	6513170	58.215 ug/ml
6) N-OCTADECANE	10.995	6280296	55.369 ug/ml
7) N-EICOSANE	12.307	6064266	53.048 ug/ml
8) N-DOCOSANE	13.506	5929360	51.870 ug/ml
10) N-TETRACOSANE	14.612	5904238	51.112 ug/ml
11) N-HEXACOSANE	15.634	5840118	50.946 ug/ml
12) N-OCTACOSANE	16.589	5871164	50.750 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG042219\
Data File : FG004764.D
Signal(s) : FID1A.ch
Acq On : 22 Apr 2019 11:29
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 22 22:16:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG032719.M
Quant Title :
QLast Update : Wed Mar 27 13:30:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

