

rteres

Instrument :

FID_G

LabSampleId :

50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/26/2022

Supervised By :mohammad ahmed 04/27/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG04252
Data File : FG009665.D
Signal(s) : FID1A.ch
Acq On : 25 Apr 2022 19:21
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040122.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.221	2.188	2.277	BB	890475	6867840	98.80%	6.007%
2	4.319	4.279	4.442	BB	629667	6940832	99.85%	6.071%
3	6.348	6.315	6.461	BB	553004	6951420	100.00%	6.080%
4	8.148	8.114	8.250	BB	512027	6900353	99.27%	6.035%
5	9.755	9.710	9.863	BB	495817	6858996	98.67%	5.999%
6	11.203	11.134	11.293	BB	506949	6928736	99.67%	6.060%
7	12.518	12.475	12.627	BB	492537	6691580	96.26%	5.853%
8	13.721	13.681	13.808	BB	484825	6573646	94.57%	5.750%
9	14.625	14.577	14.720	BB	417053	5850749	84.17%	5.117%
10	14.830	14.773	14.927	BB	481530	6530536	93.95%	5.712%
11	15.855	15.803	15.947	BB	440385	6471985	93.10%	5.661%
12	16.810	16.745	16.903	BB	435441	6454657	92.85%	5.646%
13	17.700	17.639	17.786	BB	423933	6447007	92.74%	5.639%
14	18.538	18.470	18.634	BB	394829	6342684	91.24%	5.548%
15	19.322	19.252	19.415	BB	401751	5984481	86.09%	5.234%
16	20.066	19.992	20.175	BV	351574	5552866	79.88%	4.857%
17	20.805	20.732	20.900	BB	272960	5024058	72.27%	4.394%
18	21.717	21.627	21.787	BV	205818	4958573	71.33%	4.337%
Sum of corrected areas:						114331001		

FG040122.M Tue Apr 26 17:09:12 2022