

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG042822\
 Data File : FG009688.D
 Signal(s) : FID1A.ch
 Acq On : 28 Apr 2022 10:15
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/29/2022
 Supervised By :mohammad ahmed 05/02/2022

Integration File: Sample.e
 Quant Time: Apr 29 01:40:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040122.M
 Quant Title :
 QLast Update : Fri Apr 08 10:19:35 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.622	6155869	47.784 ug/mlm
Target Compounds			
2) N-DECANE	4.317	7334730	49.401 ug/ml
3) N-DODECANE	6.345	7334444	48.665 ug/ml
4) N-TETRADECANE	8.145	7306111	47.696 ug/ml
5) N-HEXADECANE	9.752	7248333	46.970 ug/ml
6) N-OCTADECANE	11.201	7324991	46.556 ug/ml
7) N-EICOSANE	12.514	7099071	46.872 ug/ml
8) N-DOCOSANE	13.718	6977209	47.278 ug/ml
10) N-TETRACOSANE	14.828	6907836	48.187 ug/ml
11) N-HEXACOSANE	15.853	6836336	49.766 ug/ml
12) N-OCTACOSANE	16.807	6826752	52.308 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG042822\
Data File : FG009688.D
Signal(s) : FID1A.ch
Acq On : 28 Apr 2022 10:15
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/29/2022
Supervised By :mohammad ahmed 05/02/2022

Integration File: Sample.e
Quant Time: Apr 29 01:40:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040122.M
Quant Title :
QLast Update : Fri Apr 08 10:19:35 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

