

rteres

Instrument :

FID_G

LabSampleId :

50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/29/2022

Supervised By :mohammad ahmed 05/02/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG04282
Data File : FG009692.D
Signal(s) : FID1A.ch
Acq On : 28 Apr 2022 12:20
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040122.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	4.317	4.278	4.436	BB	662854	7299040	99.96%	9.451%
2	6.345	6.308	6.462	BB	576594	7302185	100.00%	9.455%
3	8.146	8.110	8.261	BB	530134	7266690	99.51%	9.409%
4	9.753	9.710	9.873	BB	514075	7225900	98.96%	9.357%
5	11.200	11.164	11.296	BB	529509	7291529	99.85%	9.442%
6	12.516	12.468	12.644	BB	503593	7084627	97.02%	9.174%
7	13.721	13.643	13.814	BB	500844	6960668	95.32%	9.013%
8	14.623	14.576	14.725	BB	432020	6228368	85.29%	8.065%
9	14.827	14.773	14.937	BB	469591	6905001	94.56%	8.941%
10	15.854	15.797	15.962	BB	479639	6832193	93.56%	8.847%
11	16.807	16.747	16.912	BB	473178	6831954	93.56%	8.846%
Sum of corrected areas:						77228155		

FG040122.M Fri Apr 29 14:46:21 2022