

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD

rteres

Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG041124.D
Data File : FG013733.D
Signal(s) : FID1A.ch
Acq On : 01 May 2024 01:57
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Reviewed By :Yogesh Patel 05/01/2024
Supervised By :mohammad ahmed 05/02/2024

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG041124.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.058	2.018	2.131	BB	522727	6145026	92.17%	5.673%
2	4.585	4.548	4.630	BB	625932	6420054	96.29%	5.927%
3	6.757	6.720	6.795	BB	639485	6526199	97.88%	6.025%
4	8.591	8.554	8.628	BB	617430	6558787	98.37%	6.055%
5	10.204	10.167	10.253	BB	587959	6568308	98.52%	6.064%
6	11.653	11.610	11.694	BB	582915	6659072	99.88%	6.148%
7	12.969	12.916	13.014	BB	546430	6667243	100.00%	6.155%
8	14.171	14.134	14.209	VB	525689	6381701	95.72%	5.892%
9	15.074	15.020	15.107	BV	398394	5490369	82.35%	5.069%
10	15.279	15.198	15.332	BB	485168	6273505	94.09%	5.792%
11	16.304	16.250	16.358	BB	465405	6114382	91.71%	5.645%
12	17.255	17.181	17.312	BB	439691	5950740	89.25%	5.494%
13	18.147	18.084	18.188	BB	420092	5933463	88.99%	5.478%
14	18.981	18.920	19.021	BB	415624	5823648	87.35%	5.376%
15	19.769	19.711	19.818	BB	374908	5537607	83.06%	5.112%
16	20.510	20.456	20.575	BV	341727	5271312	79.06%	4.866%
17	21.264	21.181	21.358	BB	269305	4961266	74.41%	4.580%
18	22.213	22.113	22.331	BV	199844	5037039	75.55%	4.650%

Sum of corrected areas: 108319719

FG041124.M Wed May 01 05:32:16 2024