

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG050421\
 Data File : FG007969.D
 Signal(s) : FID1A.ch
 Acq On : 04 May 2021 15:40
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 05 01:36:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG042121.M
 Quant Title :
 QLast Update : Thu Apr 22 01:54:01 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.793	6939227	52.124 ug/ml
Target Compounds			
2) N-DECANE	4.486	9322876	61.673 ug/ml
3) N-DODECANE	6.516	9246966	60.440 ug/ml
4) N-TETRADECANE	8.319	9188706	58.869 ug/ml
5) N-HEXADECANE	9.925	8898974	57.435 ug/ml
6) N-OCTADECANE	11.373	8807350	56.014 ug/ml
7) N-EICOSANE	12.687	8438009	54.453 ug/ml
8) N-DOCOSANE	13.890	8113978	52.802 ug/ml
10) N-TETRACOSANE	14.998	7811518	51.338 ug/ml
11) N-HEXACOSANE	16.021	7511066	50.427 ug/ml
12) N-OCTACOSANE	16.975	7378244	50.383 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG050421\
Data File : FG007969.D
Signal(s) : FID1A.ch
Acq On : 04 May 2021 15:40
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 05 01:36:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG042121.M
Quant Title :
QLast Update : Thu Apr 22 01:54:01 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

