

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG050919\
 Data File : FG004906.D
 Signal(s) : FID1A.ch
 Acq On : 09 May 2019 13:34
 Operator : AJ\SJ
 Sample : 20 TRPH STD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 20 TRPH STD

Manual Integrations
 APPROVED

Ankita
 5/10/2019 12:21:07 PM

Integration File: autoint1.e
 Quant Time: May 09 14:00:25 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050919.M
 Quant Title :
 QLast Update : Thu May 09 13:59:41 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.784	2367803	19.810 ug/ml
Target Compounds			
1) N-OCTANE	2.371	2720911	19.631 ug/ml
2) N-DECANE	4.484	2787146	19.672 ug/ml
3) N-DODECANE	6.513	2825808	19.676 ug/ml
4) N-TETRADECANE	8.316	2822616	19.676 ug/ml
5) N-HEXADECANE	9.922	2840618	19.678 ug/ml
6) N-OCTADECANE	11.369	2799363	19.656 ug/ml
7) N-EICOSANE	12.682	2785868	19.670 ug/ml
8) N-DOCOSANE	13.883	2728107	19.728 ug/ml
10) N-TETRACOSANE	14.988	2672968	19.784 ug/ml
11) N-HEXACOSANE	16.011	2596263	19.835 ug/ml
12) N-OCTACOSANE	16.962	2510181	19.972 ug/ml
13) N-TRIACONTANE	17.850	2468612	20.233 ug/ml
14) N-DOTRIACONTANE	18.685	2374636	20.435 ug/ml
15) N-TETRATRIACONTANE	19.471	2261224	20.550 ug/mlm
16) N-HEXATRIACONTANE	20.215	2118911	20.043 ug/ml
17) N-OCTATRIACONTANE	20.978	1922094	19.749 ug/mlm
18) N-TETRACONTANE	21.934	1956471	20.188 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG050919\
Data File : FG004906.D
Signal(s) : FID1A.ch
Acq On : 09 May 2019 13:34
Operator : AJ\SJ
Sample : 20 TRPH STD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
FID_G
ClientSampled :
20 TRPH STD

Manual Integrations
APPROVED

Ankita
5/10/2019 12:21:07 PM

Integration File: autoint1.e
Quant Time: May 09 14:00:25 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050919.M
Quant Title :
QLast Update : Thu May 09 13:59:41 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

