

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG051019\
 Data File : FG004935.D
 Signal(s) : FID1A.ch
 Acq On : 10 May 2019 16:42
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 10 22:48:23 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050919.M
 Quant Title :
 QLast Update : Thu May 09 14:52:33 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.790	6175082	51.456 ug/ml
Target Compounds			
1) N-OCTANE	2.369	7387379	53.419 ug/ml
2) N-DECANE	4.486	7594943	53.734 ug/ml
3) N-DODECANE	6.517	7637602	53.236 ug/ml
4) N-TETRADECANE	8.320	7564830	52.715 ug/ml
5) N-HEXADECANE	9.927	7546541	52.244 ug/ml
6) N-OCTADECANE	11.373	7396166	51.876 ug/ml
7) N-EICOSANE	12.686	7319445	51.646 ug/ml
8) N-DOCOSANE	13.888	7125597	51.416 ug/ml
10) N-TETRACOSANE	14.993	6951943	51.316 ug/ml
11) N-HEXACOSANE	16.016	6730144	51.251 ug/ml
12) N-OCTACOSANE	16.969	6502752	51.154 ug/ml
13) N-TRIACONTANE	17.858	6373376	50.716 ug/ml
14) N-DOTRIACONTANE	18.695	6150176	51.325 ug/ml
15) N-TETRATRIACONTANE	19.482	5995897	53.276 ug/mlm
16) N-HEXATRIACONTANE	20.226	5622557	53.000 ug/ml
17) N-OCTATRIACONTANE	20.994	5263606	54.320 ug/ml
18) N-TETRACONTANE	21.957	5339989	54.825 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG051019\
Data File : FG004935.D
Signal(s) : FID1A.ch
Acq On : 10 May 2019 16:42
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 10 22:48:23 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050919.M
Quant Title :
QLast Update : Thu May 09 14:52:33 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

