

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG051221\
 Data File : FG008075.D
 Signal(s) : FID1A.ch
 Acq On : 12 May 2021 15:12
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 13 01:10:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG042121.M
 Quant Title :
 QLast Update : Thu Apr 22 01:54:01 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.786	6876488	51.653 ug/ml
Target Compounds			
1) N-OCTANE	2.368	8691037	58.602 ug/ml
2) N-DECANE	4.483	9052381	59.884 ug/ml
3) N-DODECANE	6.513	8952610	58.516 ug/ml
4) N-TETRADECANE	8.315	8853962	56.724 ug/ml
5) N-HEXADECANE	9.921	8553409	55.205 ug/ml
6) N-OCTADECANE	11.368	8483914	53.957 ug/ml
7) N-EICOSANE	12.682	8171798	52.735 ug/ml
8) N-DOCOSANE	13.885	7927527	51.589 ug/ml
10) N-TETRACOSANE	14.992	7709053	50.664 ug/ml
11) N-HEXACOSANE	16.015	7485505	50.255 ug/ml
12) N-OCTACOSANE	16.968	7392905	50.483 ug/ml
13) N-TRIACONTANE	17.858	7339384	51.347 ug/ml
14) N-DOTRIACONTANE	18.693	7237631	53.053 ug/ml
15) N-TETRATRIACONTANE	19.478	7153707	55.655 ug/ml
16) N-HEXATRIACONTANE	20.220	6888862	58.867 ug/ml
17) N-OCTATRIACONTANE	20.984	6509093	60.110 ug/ml
18) N-TETRACONTANE	21.946	6347731	57.947 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG051221\
Data File : FG008075.D
Signal(s) : FID1A.ch
Acq On : 12 May 2021 15:12
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 13 01:10:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG042121.M
Quant Title :
QLast Update : Thu Apr 22 01:54:01 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

