

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG051520\
 Data File : FG006875.D
 Signal(s) : FID1A.ch
 Acq On : 15 May 2020 21:50
 Operator : DD\AJ
 Sample : FG051520ICV
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 16 01:43:58 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG051520.M
 Quant Title :
 QLast Update : Sat May 16 01:41:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.795	6349475	50.746 ug/ml
Target Compounds			
1) N-OCTANE	2.374	6984206	51.134 ug/ml
2) N-DECANE	4.494	7152988	51.359 ug/ml
3) N-DODECANE	6.523	7323557	51.536 ug/ml
4) N-TETRADECANE	8.324	7406096	51.595 ug/ml
5) N-HEXADECANE	9.930	7275362	51.409 ug/ml
6) N-OCTADECANE	11.375	7492325	51.153 ug/ml
7) N-EICOSANE	12.690	7309712	50.975 ug/ml
8) N-DOCOSANE	13.892	7326503	50.754 ug/ml
10) N-TETRACOSANE	14.999	7234608	50.541 ug/ml
11) N-HEXACOSANE	16.023	7153775	50.224 ug/ml
12) N-OCTACOSANE	16.976	7079335	49.606 ug/ml
13) N-TRIACONTANE	17.868	7060052	49.035 ug/ml
14) N-DOTRIACONTANE	18.704	6770955	49.031 ug/ml
15) N-TETRATRIACONTANE	19.489	6470169	49.481 ug/ml
16) N-HEXATRIACONTANE	20.232	5979916	49.547 ug/ml
17) N-OCTATRIACONTANE	21.003	5468017	50.003 ug/ml
18) N-TETRACONTANE	21.972	5467047	50.663 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG051520\
Data File : FG006875.D
Signal(s) : FID1A.ch
Acq On : 15 May 2020 21:50
Operator : DD\AJ
Sample : FG051520ICV
Misc :
ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 16 01:43:58 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG051520.M
Quant Title :
QLast Update : Sat May 16 01:41:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

