

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG051719\
 Data File : FG004967.D
 Signal(s) : FID1A.ch
 Acq On : 17 May 2019 13:41
 Operator : AJ\SJ
 Sample : PB119814BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 PB119814BS

Manual Integrations
 APPROVED

Ankita
 5/20/2019 11:42:27 AM

Integration File: autoint1.e
 Quant Time: May 18 04:50:07 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050919.M
 Quant Title :
 QLast Update : Thu May 09 14:52:33 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.783	2032744	16.939 ug/ml
Target Compounds			
1) N-OCTANE	2.368	2664052	19.264 ug/ml
2) N-DECANE	4.483	2708366	19.162 ug/ml
3) N-DODECANE	6.512	2710154	18.891 ug/ml
4) N-TETRADECANE	8.314	2658294	18.524 ug/ml
5) N-HEXADECANE	9.921	2623179	18.160 ug/ml
6) N-OCTADECANE	11.367	2536527	17.791 ug/ml
7) N-EICOSANE	12.681	2480501	17.502 ug/ml
8) N-DOCOSANE	13.882	2372911	17.122 ug/ml
10) N-TETRACOSANE	14.988	2288814	16.895 ug/ml
11) N-HEXACOSANE	16.011	2175133	16.564 ug/ml
12) N-OCTACOSANE	16.962	2094306	16.475 ug/ml
13) N-TRIACONTANE	17.852	2062543	16.412 ug/ml
14) N-DOTRIACONTANE	18.688	2067134	17.251 ug/ml
15) N-TETRATRIACONTANE	19.472	2007386	17.836 ug/mlm
16) N-HEXATRIACONTANE	20.216	1919242	18.092 ug/ml
17) N-OCTATRIACONTANE	20.982	1819089	18.773 ug/ml
18) N-TETRACONTANE	21.944	1827428	18.762 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG051719\
Data File : FG004967.D
Signal(s) : FID1A.ch
Acq On : 17 May 2019 13:41
Operator : AJ\SJ
Sample : PB119814BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
PB119814BS

Manual Integrations
APPROVED

Ankita
5/20/2019 11:42:27 AM

Integration File: autoint1.e
Quant Time: May 18 04:50:07 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050919.M
Quant Title :
QLast Update : Thu May 09 14:52:33 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

