

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG052019\
Data File : FG004985.D
Signal(s) : FID1A.ch
Acq On : 20 May 2019 12:02
Operator : AJ\SJ
Sample : PB119860BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 21 01:07:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050919.M
Quant Title :
QLast Update : Thu May 09 14:52:33 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.784	2014891	16.790 ug/ml
Target Compounds				
12)	N-OCTACOSANE	16.960	25107	0.198 ug/ml
13)	N-TRIACONTANE	17.849	30330	0.241 ug/ml
14)	N-DOTRIACONTANE	18.686	29217	0.244 ug/ml
16)	N-HEXATRIACONTANE	20.213	17358	0.164 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG052019\
Data File : FG004985.D
Signal(s) : FID1A.ch
Acq On : 20 May 2019 12:02
Operator : AJ\SJ
Sample : PB119860BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 21 01:07:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050919.M
Quant Title :
QLast Update : Thu May 09 14:52:33 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.784 min
Delta R.T.: -0.003 min
Response: 2014891
Conc: 16.79 ug/ml




#12 N-OCTACOSANE

R.T.: 16.960 min
Delta R.T.: -0.004 min
Response: 25107
Conc: 0.20 ug/ml




#13 N-TRIACONTANE

R.T.: 17.849 min
Delta R.T.: -0.003 min
Response: 30330
Conc: 0.24 ug/ml



#14 N-DOTRIACONTANE

R.T.: 18.686 min
Delta R.T.: -0.003 min
Response: 29217
Conc: 0.24 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.213 min
Delta R.T.: -0.005 min
Response: 17358
Conc: 0.16 ug/ml