

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG052020\
 Data File : FG006878.D
 Signal(s) : FID1A.ch
 Acq On : 20 May 2020 10:32
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: May 21 05:27:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG051520.M
 Quant Title :
 QLast Update : Sat May 16 01:41:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.792	6339938	50.670 ug/ml
Target Compounds			
2) N-DECANE	4.492	7134376	51.225 ug/ml
3) N-DODECANE	6.520	7251072	51.026 ug/ml
4) N-TETRADECANE	8.322	7297612	50.839 ug/ml
5) N-HEXADECANE	9.927	7166602	50.640 ug/ml
6) N-OCTADECANE	11.373	7411861	50.604 ug/ml
7) N-EICOSANE	12.687	7262187	50.643 ug/ml
8) N-DOCOSANE	13.889	7305788	50.610 ug/ml
10) N-TETRACOSANE	14.995	7233524	50.533 ug/ml
11) N-HEXACOSANE	16.020	7167154	50.318 ug/ml
12) N-OCTACOSANE	16.972	7106391	49.795 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG052020\
Data File : FG006878.D
Signal(s) : FID1A.ch
Acq On : 20 May 2020 10:32
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: May 21 05:27:37 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG051520.M
Quant Title :
QLast Update : Sat May 16 01:41:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

