

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG052523\
 Data File : FG012016.D
 Signal(s) : FID1A.ch
 Acq On : 25 May 2023 11:06
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: May 26 05:57:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050523.M
 Quant Title :
 QLast Update : Fri May 05 17:29:57 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.027	7886411	58.222 ug/ml
Target Compounds			
2) N-DECANE	4.550	6488795	47.305 ug/ml
3) N-DODECANE	6.722	6627899	47.460 ug/ml
4) N-TETRADECANE	8.553	6616795	47.372 ug/ml
5) N-HEXADECANE	10.164	6761629	46.239 ug/ml
6) N-OCTADECANE	11.609	6738363	44.880 ug/ml
7) N-EICOSANE	12.922	6613446	43.533 ug/ml
8) N-DOCOSANE	14.123	6269832	42.098 ug/ml
10) N-TETRACOSANE	15.229	6111221	40.778 ug/ml
11) N-HEXACOSANE	16.253	5883823	39.976 ug/ml
12) N-OCTACOSANE	17.205	5712536	39.553 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG052523\
Data File : FG012016.D
Signal(s) : FID1A.ch
Acq On : 25 May 2023 11:06
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: May 26 05:57:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050523.M
Quant Title :
QLast Update : Fri May 05 17:29:57 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

