

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG052623\
 Data File : FG012028.D
 Signal(s) : FID1A.ch
 Acq On : 26 May 2023 16:57
 Operator : YP\AJ
 Sample : 02945-05MS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 WC17MS

Integration File: autoint1.e
 Quant Time: May 27 03:43:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050523.M
 Quant Title :
 QLast Update : Fri May 05 17:29:57 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.019	1583780	11.692 ug/ml
Target Compounds			
2) N-DECANE	4.549	2325243	16.952 ug/ml
3) N-DODECANE	6.719	2394676	17.147 ug/ml
4) N-TETRADECANE	8.549	2239494	16.033 ug/ml
5) N-HEXADECANE	10.160	2508446	17.154 ug/ml
6) N-OCTADECANE	11.606	2575866	17.156 ug/ml
7) N-EICOSANE	12.918	2514372	16.551 ug/ml
8) N-DOCOSANE	14.118	2461721	16.529 ug/ml
10) N-TETRACOSANE	15.224	2576984	17.195 ug/ml
11) N-HEXACOSANE	16.248	2327307	15.812 ug/ml
12) N-OCTACOSANE	17.200	2289699	15.854 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG052623\
Data File : FG012028.D
Signal(s) : FID1A.ch
Acq On : 26 May 2023 16:57
Operator : YP\AJ
Sample : 02945-05MS
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
WC17MS

Integration File: autoint1.e
Quant Time: May 27 03:43:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050523.M
Quant Title :
QLast Update : Fri May 05 17:29:57 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

