

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG052819\
 Data File : FG005045.D
 Signal(s) : FID1A.ch
 Acq On : 28 May 2019 16:28
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 29 00:54:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050919.M
 Quant Title :
 QLast Update : Thu May 09 14:52:33 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.799	4972595	41.436 ug/ml
Target Compounds			
2) N-DECANE	4.489	6879715	48.674 ug/ml
3) N-DODECANE	6.522	6783955	47.286 ug/ml
4) N-TETRADECANE	8.327	6570161	45.783 ug/ml
5) N-HEXADECANE	9.935	6422177	44.460 ug/ml
6) N-OCTADECANE	11.382	6161736	43.218 ug/ml
7) N-EICOSANE	12.696	5998807	42.328 ug/ml
8) N-DOCOSANE	13.898	5770536	41.638 ug/ml
10) N-TETRACOSANE	15.005	5744150	42.400 ug/ml
11) N-HEXACOSANE	16.028	5500731	41.889 ug/ml
12) N-OCTACOSANE	16.980	5420675	42.642 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG052819\
Data File : FG005045.D
Signal(s) : FID1A.ch
Acq On : 28 May 2019 16:28
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 29 00:54:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050919.M
Quant Title :
QLast Update : Thu May 09 14:52:33 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

