

Method Path : P:\HPCHEM1\FID_G\Method\
 Method File : FG060118.M
 Title :
 Last Update : Fri Jun 01 15:28:31 2018
 Response Via : Initial Calibration

Calibration Files

50 =FG001088.D 100 =FG001087.D 20 =FG001089.D
 10 =FG001090.D 5 =FG001091.D

	Compound	50	100	20	10	5	Avg	%RSD
1)	N-OCTANE	1.303	1.449	1.292	1.341	1.394	1.356 E5	4.84
2)	N-DECANE	1.318	1.471	1.306	1.353	1.407	1.371 E5	4.98
3)	N-DODECANE	1.333	1.486	1.322	1.368	1.422	1.386 E5	4.91
4)	N-TETRADECANE	1.344	1.498	1.334	1.377	1.434	1.397 E5	4.89
5)	N-HEXADECANE	1.354	1.508	1.344	1.387	1.444	1.407 E5	4.88
6)	N-OCTADECANE	1.362	1.517	1.353	1.397	1.459	1.418 E5	4.90
7)	N-EICOSANE	1.352	1.507	1.344	1.386	1.451	1.408 E5	4.95
8)	N-DOCOSANE	1.345	1.500	1.338	1.380	1.444	1.401 E5	4.95
9) S	TETRACOSANE-d50 (S...	1.175	1.310	1.168	1.202	1.258	1.222 E5	4.94
10)	N-TETRACOSANE	1.333	1.490	1.327	1.368	1.434	1.390 E5	5.05
11)	N-HEXACOSANE	1.327	1.486	1.321	1.360	1.434	1.385 E5	5.19
12)	N-OCTACOSANE	1.317	1.479	1.313	1.363	1.426	1.380 E5	5.21
13)	N-TRIACONTANE	1.307	1.472	1.307	1.346	1.428	1.372 E5	5.45
14)	N-DOTRIACONTANE	1.271	1.434	1.274	1.317	1.393	1.337 E5	5.44
15)	N-TETRATRIACONTANE	1.256	1.420	1.255	1.293	1.363	1.317 E5	5.47
16)	N-HEXATRIACONTANE	1.203	1.361	1.200	1.228	1.292	1.257 E5	5.47
17)	N-OCTATRIACONTANE	1.086	1.229	1.079	1.103	1.171	1.133 E5	5.68
18)	N-TETRACONTANE	1.070	1.210	1.071	1.096	1.156	1.121 E5	5.45

(#) = Out of Range