

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060122\
 Data File : FG009798.D
 Signal(s) : FID1A.ch
 Acq On : 01 Jun 2022 12:26
 Operator : AJ\MA
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 5 TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/01/2022
 Supervised By :Ankita Jodhani 06/02/2022

Integration File: Sample.e
 Quant Time: Jun 01 13:05:49 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060122.M
 Quant Title :
 QLast Update : Wed Jun 01 12:30:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.843	519429	4.919 ug/ml
Target Compounds			
1) N-OCTANE	1.861	620720	4.942 ug/ml
2) N-DECANE	4.368	627956	4.966 ug/ml
3) N-DODECANE	6.543	630798	5.000 ug/ml
4) N-TETRADECANE	8.373	640886	5.055 ug/ml
5) N-HEXADECANE	9.983	635653	5.050 ug/ml
6) N-OCTADECANE	11.427	636587	5.055 ug/ml
7) N-EICOSANE	12.739	611077	4.996 ug/ml
8) N-DOCOSANE	13.940	603729	5.009 ug/ml
10) N-TETRACOSANE	15.046	578691	4.922 ug/ml
11) N-HEXACOSANE	16.070	556035	4.877 ug/ml
12) N-OCTACOSANE	17.022	645200	5.824 ug/ml
13) N-TRIACONTANE	17.912	511259	4.689 ug/ml
14) N-DOTRIACONTANE	18.748	508943	4.796 ug/ml
15) N-TETRATRIACONTANE	19.534	434619	4.422 ug/ml
16) N-HEXATRIACONTANE	20.278	336350	3.994 ug/ml
17) N-OCTATRIACONTANE	20.996	312321	4.212 ug/ml
18) N-TETRACONTANE	21.861	335391	4.723 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060122\
Data File : FG009798.D
Signal(s) : FID1A.ch
Acq On : 01 Jun 2022 12:26
Operator : AJ\MA
Sample : 5 TRPH STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
5 TRPH STD

Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 06/01/2022
Supervised By : Ankita Jodhani 06/02/2022

Integration File: Sample.e
Quant Time: Jun 01 13:05:49 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060122.M
Quant Title :
QLast Update : Wed Jun 01 12:30:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

