

rteres

Instrument :

FID_G

LabSampleId :

5 TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/01/2022

Supervised By :Ankita Jodhani 06/02/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060122
Data File : FG009798.D
Signal(s) : FID1A.ch
Acq On : 01 Jun 2022 12:26
Sample : 5 TRPH STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060122.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	1.861	1.821	1.930	BB	56049	620720	96.21%	6.419%
2	4.368	4.330	4.411	BB	59460	627956	97.33%	6.494%
3	6.543	6.503	6.584	BB	60949	630798	97.77%	6.523%
4	8.373	8.339	8.410	BB	60463	640886	99.33%	6.628%
5	9.983	9.953	10.028	BB	57156	635653	98.52%	6.574%
6	11.427	11.396	11.472	BB	53222	636587	98.67%	6.583%
7	12.739	12.706	12.773	BB	50100	611077	94.71%	6.319%
8	13.940	13.908	13.991	BB	45202	603729	93.57%	6.243%
9	14.843	14.803	14.894	BB	37162	519429	80.51%	5.372%
10	15.046	15.001	15.100	BB	42551	578691	89.69%	5.985%
11	16.070	16.032	16.128	BB	39196	556035	86.18%	5.750%
12	17.022	16.972	17.058	BV	38131	645200	100.00%	6.672%
13	17.912	17.863	17.982	BB	32793	511259	79.24%	5.287%
14	18.748	18.698	18.833	BB	29798	508943	78.88%	5.263%
15	19.534	19.482	19.599	BB	25689	434619	67.36%	4.495%
16	20.278	20.228	20.323	BV	21426	336350	52.13%	3.478%
17	20.996	20.944	21.067	BB	15189	312321	48.41%	3.230%
18	21.861	21.795	21.928	BB	10391	259543	40.23%	2.684%
Sum of corrected areas:						9669797		

FG060122.M Fri Jun 03 16:45:45 2022