

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060519\
Data File : FG005099.D
Signal(s) : FID1A.ch
Acq On : 05 Jun 2019 12:09
Operator : DD\AJ
Sample : RT MARKER
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 06 00:50:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050919.M
Quant Title :
QLast Update : Thu May 09 14:52:33 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
Target Compounds			
3) N-DODECANE	6.513	3439074	23.971 ug/ml
10) N-TETRACOSANE	14.976	85335	0.630 ug/ml
11) N-HEXACOSANE	16.010	32753	0.249 ug/ml
12) N-OCTACOSANE	16.963	2944632	23.164 ug/ml
13) N-TRIACONTANE	17.850	163870	1.304 ug/ml
14) N-DOTRIACONTANE	18.685	184050	1.536 ug/ml
15) N-TETRATRIACONTANE	19.472	137317	1.220 ug/ml
16) N-HEXATRIACONTANE	20.212	93766	0.884 ug/ml
17) N-OCTATRIACONTANE	20.976	57457	0.593 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060519\
Data File : FG005099.D
Signal(s) : FID1A.ch
Acq On : 05 Jun 2019 12:09
Operator : DD\AJ
Sample : RT MARKER
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 06 00:50:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050919.M
Quant Title :
QLast Update : Thu May 09 14:52:33 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#3 N-DODECANE

R.T.: 6.513 min
Delta R.T.: -0.003 min
Response: 3439074
Conc: 23.97 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 0.000 min
Exp R.T. : 14.786 min
Response: 0
Conc: N.D.




#10 N-TETRACOSANE

R.T.: 14.976 min
Delta R.T.: -0.017 min
Response: 85335
Conc: 0.63 ug/ml




#11 N-HEXACOSANE

R.T.: 16.010 min
Delta R.T.: -0.004 min
Response: 32753
Conc: 0.25 ug/ml




#12 N-OCTACOSANE

R.T.: 16.963 min
Delta R.T.: -0.001 min
Response: 2944632
Conc: 23.16 ug/ml




#13 N-TRIACONTANE

R.T.: 17.850 min
Delta R.T.: -0.003 min
Response: 163870
Conc: 1.30 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.685 min
Delta R.T.: -0.004 min
Response: 184050
Conc: 1.54 ug/ml




#15 N-TETRATRIACONTANE

R.T.: 19.472 min
Delta R.T.: -0.002 min
Response: 137317
Conc: 1.22 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.212 min
Delta R.T.: -0.005 min
Response: 93766
Conc: 0.88 ug/ml



#17 N-OCTATRIACONTANE

R.T.: 20.976 min
Delta R.T.: -0.004 min
Response: 57457
Conc: 0.59 ug/ml