

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060519\
 Data File : FG005098.D
 Signal(s) : FID1A.ch
 Acq On : 05 Jun 2019 11:40
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Jun 06 00:50:36 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050919.M
 Quant Title :
 QLast Update : Thu May 09 14:52:33 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.788	5827238	48.558 ug/ml
Target Compounds			
2) N-DECANE	4.484	8186518	57.919 ug/ml
3) N-DODECANE	6.515	7982392	55.640 ug/ml
4) N-TETRADECANE	8.318	7663295	53.401 ug/ml
5) N-HEXADECANE	9.926	7434204	51.466 ug/ml
6) N-OCTADECANE	11.372	7111209	49.877 ug/ml
7) N-EICOSANE	12.686	6931682	48.910 ug/ml
8) N-DOCOSANE	13.887	6718076	48.476 ug/ml
10) N-TETRACOSANE	14.992	6695331	49.422 ug/ml
11) N-HEXACOSANE	16.015	6564352	49.989 ug/ml
12) N-OCTACOSANE	16.967	6640205	52.235 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060519\
Data File : FG005098.D
Signal(s) : FID1A.ch
Acq On : 05 Jun 2019 11:40
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Jun 06 00:50:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG050919.M
Quant Title :
QLast Update : Thu May 09 14:52:33 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

