

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060520\
 Data File : FG006935.D
 Signal(s) : FID1A.ch
 Acq On : 05 Jun 2020 16:09
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 06 01:52:25 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG051520.M
 Quant Title :
 QLast Update : Sat May 16 01:41:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.789	6078987	48.584 ug/ml
Target Compounds			
1) N-OCTANE	2.377	6735837	49.316 ug/ml
2) N-DECANE	4.492	7381660	53.001 ug/ml
3) N-DODECANE	6.521	7460032	52.497 ug/ml
4) N-TETRADECANE	8.320	7445070	51.867 ug/ml
5) N-HEXADECANE	9.926	7259352	51.296 ug/ml
6) N-OCTADECANE	11.372	7429812	50.727 ug/ml
7) N-EICOSANE	12.686	7179851	50.069 ug/ml
8) N-DOCOSANE	13.887	7081097	49.054 ug/ml
10) N-TETRACOSANE	14.993	6840204	47.785 ug/ml
11) N-HEXACOSANE	16.016	6605080	46.372 ug/ml
12) N-OCTACOSANE	16.969	6439461	45.122 ug/ml
13) N-TRIACONTANE	17.858	6442489	44.746 ug/ml
14) N-DOTRIACONTANE	18.692	6126793	44.366 ug/ml
15) N-TETRATRIACONTANE	19.478	5742318	43.915 ug/ml
16) N-HEXATRIACONTANE	20.222	5304220	43.948 ug/ml
17) N-OCTATRIACONTANE	20.989	4865889	44.497 ug/ml
18) N-TETRACONTANE	21.957	4920380	45.597 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060520\
Data File : FG006935.D
Signal(s) : FID1A.ch
Acq On : 05 Jun 2020 16:09
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 06 01:52:25 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG051520.M
Quant Title :
QLast Update : Sat May 16 01:41:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

