

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060520\
 Data File : FG060520.D
 Signal(s) : FID1A.ch
 Acq On : 06 Jun 2020 00:32
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 06 01:55:10 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG051520.M
 Quant Title :
 QLast Update : Sat May 16 01:41:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.785	5829276	46.588 ug/ml
Target Compounds			
1) N-OCTANE	2.376	6254994	45.795 ug/ml
2) N-DECANE	4.493	6875177	49.364 ug/ml
3) N-DODECANE	6.521	6960543	48.982 ug/ml
4) N-TETRADECANE	8.320	6964057	48.516 ug/ml
5) N-HEXADECANE	9.925	6797701	48.034 ug/ml
6) N-OCTADECANE	11.371	6984111	47.684 ug/ml
7) N-EICOSANE	12.683	6781050	47.288 ug/ml
8) N-DOCOSANE	13.885	6739923	46.690 ug/ml
10) N-TETRACOSANE	14.991	6570022	45.898 ug/ml
11) N-HEXACOSANE	16.015	6400325	44.935 ug/ml
12) N-OCTACOSANE	16.966	6290554	44.079 ug/ml
13) N-TRIACONTANE	17.855	6336912	44.012 ug/ml
14) N-DOTRIACONTANE	18.689	6054404	43.842 ug/ml
15) N-TETRATRIACONTANE	19.474	5687146	43.493 ug/ml
16) N-HEXATRIACONTANE	20.216	5257694	43.563 ug/ml
17) N-OCTATRIACONTANE	20.980	4832412	44.191 ug/ml
18) N-TETRACONTANE	21.944	4896586	45.377 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060520\
Data File : FG006951.D
Signal(s) : FID1A.ch
Acq On : 06 Jun 2020 00:32
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 06 01:55:10 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG051520.M
Quant Title :
QLast Update : Sat May 16 01:41:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

