

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060821\
 Data File : FG008253.D
 Signal(s) : FID1A.ch
 Acq On : 08 Jun 2021 17:18
 Operator : DD\AJ
 Sample : M2637-05MSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 09 01:09:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG052521.M
 Quant Title :
 QLast Update : Wed May 26 08:20:50 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.780	2231325	18.802 ug/ml
Target Compounds				
1)	N-OCTANE	2.365	2178665	16.952 ug/ml
2)	N-DECANE	4.477	2637296	20.129 ug/ml
3)	N-DODECANE	6.506	2700840	20.245 ug/ml
4)	N-TETRADECANE	8.308	2744746	20.132 ug/ml
5)	N-HEXADECANE	9.915	2668796	19.602 ug/ml
6)	N-OCTADECANE	11.361	2635181	19.020 ug/ml
7)	N-EICOSANE	12.676	2612707	19.089 ug/ml
8)	N-DOCOSANE	13.878	2546871	18.764 ug/ml
10)	N-TETRACOSANE	14.984	2525052	18.770 ug/ml
11)	N-HEXACOSANE	16.008	2466678	18.768 ug/ml
12)	N-OCTACOSANE	16.961	2475914	19.102 ug/ml
13)	N-TRIACONTANE	17.849	2446961	19.168 ug/ml
14)	N-DOTRIACONTANE	18.685	2514080	20.231 ug/ml
15)	N-TETRATRIACONTANE	19.469	2337717	19.984 ug/ml
16)	N-HEXATRIACONTANE	20.210	2290476	20.682 ug/ml
17)	N-OCTATRIACONTANE	20.972	2174377	21.049 ug/ml
18)	N-TETRACONTANE	21.930	2164985	20.174 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060821\
Data File : FG008253.D
Signal(s) : FID1A.ch
Acq On : 08 Jun 2021 17:18
Operator : DD\AJ
Sample : M2637-05MSD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 09 01:09:47 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG052521.M
Quant Title :
QLast Update : Wed May 26 08:20:50 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

