

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060923\
 Data File : FG012140.D
 Signal(s) : FID1A.ch
 Acq On : 09 Jun 2023 14:29
 Operator : YP\AJ
 Sample : PB153320BS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 PB153320BS

Integration File: autoint1.e
 Quant Time: Jun 09 23:46:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060223.M
 Quant Title :
 QLast Update : Thu Jun 01 17:30:11 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.010	1911921	14.340 ug/ml
Target Compounds			
2) N-DECANE	4.543	2309488	15.674 ug/ml
3) N-DODECANE	6.712	2345041	15.623 ug/ml
4) N-TETRADECANE	8.541	2323062	15.505 ug/ml
5) N-HEXADECANE	10.152	2420564	15.471 ug/ml
6) N-OCTADECANE	11.597	2457182	15.451 ug/ml
7) N-EICOSANE	12.909	2374606	14.964 ug/ml
8) N-DOCOSANE	14.109	2298165	15.041 ug/ml
10) N-TETRACOSANE	15.214	2221177	14.870 ug/ml
11) N-HEXACOSANE	16.238	2149591	15.179 ug/ml
12) N-OCTACOSANE	17.189	2104348	15.545 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060923\
Data File : FG012140.D
Signal(s) : FID1A.ch
Acq On : 09 Jun 2023 14:29
Operator : YP\AJ
Sample : PB153320BS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
PB153320BS

Integration File: autoint1.e
Quant Time: Jun 09 23:46:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060223.M
Quant Title :
QLast Update : Thu Jun 01 17:30:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

