

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061021\
 Data File : FG008281.D
 Signal(s) : FID1A.ch
 Acq On : 10 Jun 2021 12:42
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 11 02:30:09 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG052521.M
 Quant Title :
 QLast Update : Wed May 26 08:20:50 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.779	5974943	50.347 ug/ml
Target Compounds			
1) N-OCTANE	2.363	7103760	55.275 ug/ml
2) N-DECANE	4.477	7385455	56.370 ug/ml
3) N-DODECANE	6.506	7385237	55.359 ug/ml
4) N-TETRADECANE	8.308	7396070	54.249 ug/ml
5) N-HEXADECANE	9.914	7204164	52.914 ug/ml
6) N-OCTADECANE	11.361	7205240	52.005 ug/ml
7) N-EICOSANE	12.676	7012495	51.234 ug/ml
8) N-DOCOSANE	13.878	6853607	50.494 ug/ml
10) N-TETRACOSANE	14.984	6732661	50.048 ug/ml
11) N-HEXACOSANE	16.009	6536816	49.735 ug/ml
12) N-OCTACOSANE	16.962	6471718	49.931 ug/ml
13) N-TRIACONTANE	17.853	6455329	50.568 ug/ml
14) N-DOTRIACONTANE	18.686	6369094	51.253 ug/ml
15) N-TETRATRIACONTANE	19.472	6265680	53.561 ug/ml
16) N-HEXATRIACONTANE	20.212	6064692	54.760 ug/ml
17) N-OCTATRIACONTANE	20.974	5715425	55.328 ug/ml
18) N-TETRACONTANE	21.933	5869976	54.699 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061021\
Data File : FG008281.D
Signal(s) : FID1A.ch
Acq On : 10 Jun 2021 12:42
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 11 02:30:09 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG052521.M
Quant Title :
QLast Update : Wed May 26 08:20:50 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

