

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061021\
 Data File : FG008283.D
 Signal(s) : FID1A.ch
 Acq On : 10 Jun 2021 15:53
 Operator : DD\AJ
 Sample : PB136978BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 11 09:18:34 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG052521.M
 Quant Title :
 QLast Update : Wed May 26 08:20:50 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.777	1903685	16.041 ug/ml
Target Compounds			
1) N-OCTANE	2.362	2151803	16.743 ug/ml
2) N-DECANE	4.474	2304042	17.586 ug/ml
3) N-DODECANE	6.503	2316347	17.363 ug/ml
4) N-TETRADECANE	8.304	2306689	16.919 ug/ml
5) N-HEXADECANE	9.910	2249824	16.525 ug/ml
6) N-OCTADECANE	11.358	2220720	16.028 ug/ml
7) N-EICOSANE	12.672	2205592	16.114 ug/ml
8) N-DOCOSANE	13.874	2146772	15.817 ug/ml
10) N-TETRACOSANE	14.983	2097601	15.593 ug/ml
11) N-HEXACOSANE	16.006	2029003	15.438 ug/ml
12) N-OCTACOSANE	16.958	2004018	15.462 ug/ml
13) N-TRIACONTANE	17.848	1991246	15.598 ug/ml
14) N-DOTRIACONTANE	18.684	1956399	15.744 ug/ml
15) N-TETRATRIACONTANE	19.469	1807027	15.447 ug/ml
16) N-HEXATRIACONTANE	20.209	1726968	15.593 ug/ml
17) N-OCTATRIACONTANE	20.973	1643615	15.911 ug/ml
18) N-TETRACONTANE	21.929	1631493	15.203 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061021\
Data File : FG008283.D
Signal(s) : FID1A.ch
Acq On : 10 Jun 2021 15:53
Operator : DD\AJ
Sample : PB136978BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 11 09:18:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG052521.M
Quant Title :
QLast Update : Wed May 26 08:20:50 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

