

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061021\
 Data File : FG008287.D
 Signal(s) : FID1A.ch
 Acq On : 10 Jun 2021 18:00
 Operator : DD\AJ
 Sample : M2638-01MS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 11 09:19:00 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG052521.M
 Quant Title :
 QLast Update : Wed May 26 08:20:50 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.778	1415665	11.929 ug/ml
Target Compounds			
1) N-OCTANE	2.363	2721461	21.176 ug/ml
2) N-DECANE	4.476	3033000	23.150 ug/ml
3) N-DODECANE	6.504	3070278	23.015 ug/ml
4) N-TETRADECANE	8.306	3091286	22.674 ug/ml
5) N-HEXADECANE	9.912	3002332	22.052 ug/ml
6) N-OCTADECANE	11.359	2982339	21.525 ug/ml
7) N-EICOSANE	12.674	2990910	21.852 ug/ml
8) N-DOCOSANE	13.876	2952183	21.750 ug/ml
10) N-TETRACOSANE	14.984	3096087	23.015 ug/ml
11) N-HEXACOSANE	16.008	2755677	20.966 ug/ml
12) N-OCTACOSANE	16.961	2781929	21.463 ug/ml
13) N-TRIACONTANE	17.849	2693844	21.102 ug/ml
14) N-DOTRIACONTANE	18.683	2712132	21.825 ug/ml
15) N-TETRATRIACONTANE	19.468	2507054	21.431 ug/ml
16) N-HEXATRIACONTANE	20.210	2457115	22.186 ug/ml
17) N-OCTATRIACONTANE	20.971	2311822	22.379 ug/ml
18) N-TETRACONTANE	21.929	2239971	20.873 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061021\
Data File : FG008287.D
Signal(s) : FID1A.ch
Acq On : 10 Jun 2021 18:00
Operator : DD\AJ
Sample : M2638-01MS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 11 09:19:00 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG052521.M
Quant Title :
QLast Update : Wed May 26 08:20:50 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

