

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061022\
 Data File : FG009845.D
 Signal(s) : FID1A.ch
 Acq On : 10 Jun 2022 14:11
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 11 00:30:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060122.M
 Quant Title :
 QLast Update : Wed Jun 01 13:06:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.835	4965897	47.179 ug/ml
Target Compounds			
1) N-OCTANE	1.852	5882127	46.941 ug/ml
2) N-DECANE	4.358	5900157	46.720 ug/ml
3) N-DODECANE	6.534	5867454	46.509 ug/ml
4) N-TETRADECANE	8.366	5858018	46.103 ug/ml
5) N-HEXADECANE	9.976	5816530	46.116 ug/ml
6) N-OCTADECANE	11.421	5819277	46.107 ug/ml
7) N-EICOSANE	12.733	5692174	46.542 ug/ml
8) N-DOCOSANE	13.933	5634441	46.727 ug/ml
10) N-TETRACOSANE	15.039	5523363	47.127 ug/ml
11) N-HEXACOSANE	16.063	5430173	47.865 ug/ml
12) N-OCTACOSANE	17.015	5348326	46.738 ug/ml
13) N-TRIACONTANE	17.906	5321670	49.421 ug/ml
14) N-DOTRIACONTANE	18.741	5136035	48.794 ug/ml
15) N-TETRATRIACONTANE	19.525	4712867	49.084 ug/ml
16) N-HEXATRIACONTANE	20.265	4092360	50.637 ug/mlm
17) N-OCTATRIACONTANE	20.983	3280701	45.683 ug/ml
18) N-TETRACONTANE	21.843	3035076	43.221 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061022\
Data File : FG009845.D
Signal(s) : FID1A.ch
Acq On : 10 Jun 2022 14:11
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 11 00:30:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060122.M
Quant Title :
QLast Update : Wed Jun 01 13:06:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

