

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061022\
 Data File : FG009854.D
 Signal(s) : FID1A.ch
 Acq On : 10 Jun 2022 19:35
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 11 01:05:58 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060122.M
 Quant Title :
 QLast Update : Wed Jun 01 13:06:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.834	4899341	46.547 ug/ml
Target Compounds			
1) N-OCTANE	1.851	5887668	46.985 ug/ml
2) N-DECANE	4.358	5890445	46.643 ug/ml
3) N-DODECANE	6.535	5822227	46.151 ug/ml
4) N-TETRADECANE	8.365	5773942	45.442 ug/ml
5) N-HEXADECANE	9.975	5730370	45.433 ug/ml
6) N-OCTADECANE	11.420	5733654	45.428 ug/ml
7) N-EICOSANE	12.733	5594542	45.744 ug/ml
8) N-DOCOSANE	13.934	5530024	45.861 ug/ml
10) N-TETRACOSANE	15.040	5428869	46.321 ug/ml
11) N-HEXACOSANE	16.063	5351227	47.169 ug/ml
12) N-OCTACOSANE	17.014	5293146	46.256 ug/ml
13) N-TRIACONTANE	17.904	5290772	49.134 ug/ml
14) N-DOTRIACONTANE	18.739	5136429	48.798 ug/ml
15) N-TETRATRIACONTANE	19.527	4727090	49.232 ug/ml
16) N-HEXATRIACONTANE	20.266	3936968	48.715 ug/ml
17) N-OCTATRIACONTANE	20.982	3223076	44.880 ug/ml
18) N-TETRACONTANE	21.841	2995301	42.654 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061022\
Data File : FG009854.D
Signal(s) : FID1A.ch
Acq On : 10 Jun 2022 19:35
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 11 01:05:58 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060122.M
Quant Title :
QLast Update : Wed Jun 01 13:06:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

