

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061423\
 Data File : FG012179.D
 Signal(s) : FID1A.ch
 Acq On : 14 Jun 2023 16:01
 Operator : YP\AJ
 Sample : 03139-01MSD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 WASTEMSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/15/2023
 Supervised By :Ankita Jodhani 06/15/2023

Integration File: autoint1.e
 Quant Time: Jun 15 02:12:04 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060223.M
 Quant Title :
 QLast Update : Thu Jun 01 17:30:11 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.003	1432991	10.748 ug/ml
Target Compounds			
1) N-OCTANE	2.033	1780966	12.806 ug/ml
2) N-DECANE	4.538	2295451	15.579 ug/ml
3) N-DODECANE	6.707	2390204	15.924 ug/ml
4) N-TETRADECANE	8.536	2482608	16.570 ug/ml
5) N-HEXADECANE	10.145	2452641	15.676 ug/ml
6) N-OCTADECANE	11.590	2833652	17.818 ug/ml
7) N-EICOSANE	12.901	2570127	16.196 ug/ml
8) N-DOCOSANE	14.101	2546930	16.669 ug/ml
10) N-TETRACOSANE	15.206	2580076	17.272 ug/ml
11) N-HEXACOSANE	16.231	2364408	16.696 ug/ml
12) N-OCTACOSANE	17.182	2288854	16.908 ug/ml
13) N-TRIACONTANE	18.072	2489873	18.792 ug/ml
14) N-DOTRIACONTANE	18.909	2974043	23.299 ug/ml
15) N-TETRATRIACONTANE	19.694	2085210	17.131 ug/ml
16) N-HEXATRIACONTANE	20.434	1353262	11.042 ug/ml
17) N-OCTATRIACONTANE	21.175	1307668	10.249 ug/ml
18) N-TETRACONTANE	22.092	1233385	9.315 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061423\
Data File : FG012179.D
Signal(s) : FID1A.ch
Acq On : 14 Jun 2023 16:01
Operator : YP\AJ
Sample : 03139-01MSD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
WASTEMSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/15/2023
Supervised By :Ankita Jodhani 06/15/2023

Integration File: autoint1.e
Quant Time: Jun 15 02:12:04 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060223.M
Quant Title :
QLast Update : Thu Jun 01 17:30:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

