

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061621\
 Data File : FG008354.D
 Signal(s) : FID1A.ch
 Acq On : 16 Jun 2021 11:55
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 17 01:08:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG052521.M
 Quant Title :
 QLast Update : Wed May 26 08:20:50 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.775	6088708	51.306 ug/ml
Target Compounds			
1) N-OCTANE	2.359	7282565	56.666 ug/ml
2) N-DECANE	4.472	7663653	58.494 ug/ml
3) N-DODECANE	6.501	7673116	57.517 ug/ml
4) N-TETRADECANE	8.303	7650419	56.114 ug/ml
5) N-HEXADECANE	9.910	7422308	54.516 ug/ml
6) N-OCTADECANE	11.357	7425675	53.596 ug/ml
7) N-EICOSANE	12.671	7202480	52.622 ug/ml
8) N-DOCOSANE	13.874	7033212	51.818 ug/ml
10) N-TETRACOSANE	14.981	6892307	51.235 ug/ml
11) N-HEXACOSANE	16.005	6641300	50.530 ug/ml
12) N-OCTACOSANE	16.956	6540720	50.463 ug/ml
13) N-TRIACONTANE	17.847	6492785	50.861 ug/ml
14) N-DOTRIACONTANE	18.682	6405228	51.544 ug/ml
15) N-TETRATRIACONTANE	19.467	6314509	53.979 ug/ml
16) N-HEXATRIACONTANE	20.209	6128471	55.336 ug/ml
17) N-OCTATRIACONTANE	20.970	5849818	56.629 ug/ml
18) N-TETRACONTANE	21.927	5944950	55.397 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061621\
Data File : FG008354.D
Signal(s) : FID1A.ch
Acq On : 16 Jun 2021 11:55
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 17 01:08:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG052521.M
Quant Title :
QLast Update : Wed May 26 08:20:50 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

