

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061625AL\
 Data File : FG0616100.D
 Signal(s) : FID1A.ch
 Acq On : 16 Jun 2025 15:01
 Operator : YP\AJ
 Sample : Q2126-07
 Misc : 4 PPM WATER LOD
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 LOD-MDL-WATER-01-QT2-2025

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/17/2025
 Supervised By :mohammad ahmed 06/18/2025

Integration File: autoint1.e
 Quant Time: Jun 17 05:01:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\Aliphatic EPH 061325.M
 Quant Title : GC Extractables
 QLast Update : Mon Jun 16 06:52:22 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	13.538	5130301	50.914 ug/ml
Spiked Amount 50.000		Recovery =	101.83%
Target Compounds			
1) T n-Nonane (C9)	3.420	303135	2.802 ug/ml
2) T n-Decane (C10)	4.679	369924	3.372 ug/ml
4) T n-Dodecane (C12)	6.857	427657	3.884 ug/ml
6) T n-Tetradecane (C14)	8.695	464988	4.234 ug/ml
7) T n-Hexadecane (C16)	10.313	447566	3.932 ug/ml
8) T n-Octadecane (C18)	11.766	481216	4.134 ug/ml
10) T n-Eicosane (C20)	13.085	518648	4.557 ug/ml
11) T n-Heneicosane (C21)	13.700	494923	4.372 ug/ml
13) T n-Docosane (C22)	14.290	511175	4.497 ug/ml
14) T n-Tetracosane (C24)	15.401	514400	4.461 ug/ml
15) T n-Hexacosane (C26)	16.428	515173	4.477 ug/ml
16) T n-Octacosane (C28)	17.384	536747	4.729 ug/ml
17) T n-Tricontane (C30)	18.276	592200	5.172 ug/ml
18) T n-Dotriacontane (C32)	19.112	599632	5.500 ug/ml
19) T n-Tetratriacontane (C34)	19.899	555946	5.619 ug/ml
20) T n-Hexatriacontane (C36)	20.642	492166	5.798 ug/ml
21) T n-Octatriacontane (C38)	21.432	442576	5.863 ug/ml
22) T n-Tetracontane (C40)	22.433	394975	5.712 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061625AL\
Data File : FG016100.D
Signal(s) : FID1A.ch
Acq On : 16 Jun 2025 15:01
Operator : YP\AJ
Sample : Q2126-07
Misc : 4 PPM WATER LOD
ALS Vial : 33 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/17/2025
Supervised By :mohammad ahmed 06/18/2025

Integration File: autoint1.e
Quant Time: Jun 17 05:01:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\Aliphatic EPH 061325.M
Quant Title : GC Extractables
QLast Update : Mon Jun 16 06:52:22 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

