

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061722\
 Data File : FG009910.D
 Signal(s) : FID1A.ch
 Acq On : 17 Jun 2022 21:43
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/20/2022
 Supervised By :Ankita Jodhani 06/20/2022

Integration File: Sample.e
 Quant Time: Jun 17 23:50:55 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060122.M
 Quant Title :
 QLast Update : Wed Jun 01 13:06:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.839	5269213	50.061 ug/ml
Target Compounds			
1) N-OCTANE	1.851	6071704	48.454 ug/ml
2) N-DECANE	4.357	6337576	50.183 ug/ml
3) N-DODECANE	6.535	6228327	49.370 ug/ml
4) N-TETRADECANE	8.366	6176462	48.610 ug/ml
5) N-HEXADECANE	9.975	6125825	48.568 ug/ml
6) N-OCTADECANE	11.422	6129215	48.562 ug/ml
7) N-EICOSANE	12.734	5994426	49.013 ug/ml
8) N-DOCOSANE	13.935	5941903	49.277 ug/ml
10) N-TETRACOSANE	15.042	5879314	50.164 ug/ml
11) N-HEXACOSANE	16.064	5878648	51.818 ug/ml
12) N-OCTACOSANE	17.019	5961797	52.099 ug/ml
13) N-TRIACONTANE	17.909	6233540	57.890 ug/ml
14) N-DOTRIACONTANE	18.744	6347194	60.301 ug/ml
15) N-TETRATRIACONTANE	19.529	6205341	64.628 ug/ml
16) N-HEXATRIACONTANE	20.271	5613891	69.464 ug/mlm
17) N-OCTATRIACONTANE	20.991	4929973	68.648 ug/ml
18) N-TETRACONTANE	21.848	4833542	68.832 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061722\
Data File : FG009910.D
Signal(s) : FID1A.ch
Acq On : 17 Jun 2022 21:43
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 06/20/2022
Supervised By : Ankita Jodhani 06/20/2022

Integration File: Sample.e
Quant Time: Jun 17 23:50:55 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060122.M
Quant Title :
QLast Update : Wed Jun 01 13:06:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

