

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061918\
 Data File : FG001367.D
 Signal(s) : FID1A.ch
 Acq On : 19 Jun 2018 11:02
 Operator : UA\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jun 20 07:53:44 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_G\Method\FG060118.M
 Quant Title :
 QLast Update : Fri Jun 01 15:28:31 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.536	5930807	48.514 ug/ml
Target Compounds			
2) N-DECANE	4.276	7425942	54.175 ug/ml
3) N-DODECANE	6.293	7396006	53.352 ug/ml
4) N-TETRADECANE	8.085	7352602	52.618 ug/ml
5) N-HEXADECANE	9.684	7299273	51.868 ug/ml
6) N-OCTADECANE	11.125	7261531	51.227 ug/ml
7) N-EICOSANE	12.436	7130954	50.640 ug/ml
8) N-DOCOSANE	13.635	7024131	50.120 ug/ml
10) N-TETRACOSANE	14.739	6920385	49.769 ug/ml
11) N-HEXACOSANE	15.763	6834333	49.335 ug/ml
12) N-OCTACOSANE	16.713	6855737	49.688 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061918\
Data File : FG001367.D
Signal(s) : FID1A.ch
Acq On : 19 Jun 2018 11:02
Operator : UA\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jun 20 07:53:44 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_G\Method\FG060118.M
Quant Title :
QLast Update : Fri Jun 01 15:28:31 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

