

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061923\
Data File : FG012213.D
Signal(s) : FID1A.ch
Acq On : 19 Jun 2023 17:41
Operator : YP\AJ
Sample : 03308-10
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jun 20 04:12:09 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060223.M
Quant Title :
QLast Update : Thu Jun 01 17:30:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.999	1106276	8.297 ug/ml
Target Compounds			
7) N-EICOSANE	12.963	271425	1.710 ug/ml
10) N-TETRACOSANE	15.187	230562	1.543 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061923\
Data File : FG012213.D
Signal(s) : FID1A.ch
Acq On : 19 Jun 2023 17:41
Operator : YP\AJ
Sample : 03308-10
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jun 20 04:12:09 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060223.M
Quant Title :
QLast Update : Thu Jun 01 17:30:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#7 N-EICOSANE


R.T.: 12.963 min
Delta R.T.: 0.039 min
Response: 271425
Conc: 1.71 ug/ml

Instrument :
FID_G
ClientSampleId :



#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.999 min
Delta R.T.: -0.026 min
Response: 1106276
Conc: 8.30 ug/ml



#10 N-TETRACOSANE

R.T.: 15.187 min
Delta R.T.: -0.044 min
Response: 230562
Conc: 1.54 ug/ml