

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061923\
 Data File : FG012209.D
 Signal(s) : FID1A.ch
 Acq On : 19 Jun 2023 15:42
 Operator : YP\AJ
 Sample : PB153546BS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 PB153546BS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/20/2023
 Supervised By :Ankita Jodhani 06/20/2023

Integration File: autoint1.e
 Quant Time: Jun 20 04:11:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060223.M
 Quant Title :
 QLast Update : Thu Jun 01 17:30:11 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.999	2063285	15.475 ug/ml
Target Compounds			
1) N-OCTANE	2.028	2207867	15.875 ug/ml
2) N-DECANE	4.534	2759674	18.730 ug/ml
3) N-DODECANE	6.703	2833399	18.876 ug/ml
4) N-TETRADECANE	8.533	2840023	18.956 ug/ml
5) N-HEXADECANE	10.142	2971279	18.990 ug/ml
6) N-OCTADECANE	11.587	3029874	19.052 ug/ml
7) N-EICOSANE	12.898	2955866	18.627 ug/ml
8) N-DOCOSANE	14.098	2883005	18.869 ug/ml
10) N-TETRACOSANE	15.203	2805799	18.783 ug/ml
11) N-HEXACOSANE	16.225	2693717	19.021 ug/ml
12) N-OCTACOSANE	17.177	2586231	19.104 ug/ml
13) N-TRIACONTANE	18.066	2533784	19.123 ug/ml
14) N-DOTRIACONTANE	18.898	2390794	18.730 ug/ml
15) N-TETRATRIACONTANE	19.683	2014754	16.553 ug/ml
16) N-HEXATRIACONTANE	20.422	1555097	12.689 ug/mlm
17) N-OCTATRIACONTANE	21.162	1345065	10.542 ug/ml
18) N-TETRACONTANE	22.077	1269240	9.585 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG061923\
Data File : FG012209.D
Signal(s) : FID1A.ch
Acq On : 19 Jun 2023 15:42
Operator : YP\AJ
Sample : PB153546BS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
PB153546BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2023
Supervised By :Ankita Jodhani 06/20/2023

Integration File: autoint1.e
Quant Time: Jun 20 04:11:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060223.M
Quant Title :
QLast Update : Thu Jun 01 17:30:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

