

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD

rteres

Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG060223.D
Data File : FG012233.D
Signal(s) : FID1A.ch
Acq On : 20 Jun 2023 06:25
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Reviewed By :Yogesh Patel 06/20/2023
Supervised By :Ankita Jodhani 06/20/2023

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060223.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.027	1.989	2.099	BB	559335	6561955	92.64%	5.715%
2	4.537	4.503	4.581	BB	672397	6859185	96.84%	5.974%
3	6.707	6.671	6.749	BB	694866	6983597	98.59%	6.083%
4	8.536	8.498	8.581	BB	652536	6931809	97.86%	6.037%
5	10.145	10.106	10.193	BB	644822	7083139	100.00%	6.169%
6	11.590	11.550	11.634	BB	612689	7060979	99.69%	6.150%
7	12.902	12.858	12.947	BB	584154	6931406	97.86%	6.037%
8	14.101	14.056	14.147	BB	533474	6559372	92.61%	5.713%
9	15.001	14.948	15.052	BB	416326	5696436	80.42%	4.961%
10	15.206	15.145	15.257	BB	502952	6552832	92.51%	5.707%
11	16.229	16.177	16.273	BB	477274	6140625	86.69%	5.348%
12	17.179	17.128	17.220	BB	459792	5987035	84.53%	5.215%
13	18.068	18.017	18.118	BB	435114	5922910	83.62%	5.159%
14	18.902	18.845	18.941	BB	400871	5792841	81.78%	5.045%
15	19.687	19.630	19.720	BB	412158	5714501	80.68%	4.977%
16	20.428	20.400	20.463	BB	390877	5537338	78.18%	4.823%
17	21.167	21.101	21.210	BB	338805	6128764	86.53%	5.338%
18	22.085	22.010	22.158	BB	266227	6369863	89.93%	5.548%

Sum of corrected areas: 114814586

FG060223.M Tue Jun 20 06:48:09 2023