

rteres

Instrument :

FID_G

LabSampleId :

50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/27/2022
Supervised By :Ankita Jodhani 06/27/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG06242
Data File : FG009950.D
Signal(s) : FID1A.ch
Acq On : 24 Jun 2022 09:36
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060122.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	1.832	1.758	1.912	BB	456248	5085911	87.06%	5.505%
2	4.343	4.307	4.393	BB	521128	5522002	94.53%	5.977%
3	6.522	6.477	6.564	BB	539687	5604093	95.94%	6.066%
4	8.353	8.315	8.398	BB	539259	5714514	97.83%	6.186%
5	9.962	9.920	10.010	BB	522359	5768345	98.75%	6.244%
6	11.407	11.372	11.443	BB	512414	5841536	100.00%	6.323%
7	12.720	12.678	12.767	BB	471585	5762237	98.64%	6.238%
8	13.921	13.883	13.964	VB	447395	5736833	98.21%	6.210%
9	14.823	14.774	14.853	BV	362642	5083905	87.03%	5.503%
10	15.028	14.988	15.078	BB	428899	5645712	96.65%	6.111%
11	16.050	16.000	16.096	BB	405600	5577091	95.47%	6.037%
12	17.002	16.950	17.057	BB	389025	5466493	93.58%	5.917%
13	17.894	17.840	17.948	BB	377689	5403163	92.50%	5.849%
14	18.727	18.668	18.780	BB	367661	5091152	87.15%	5.511%
15	19.513	19.453	19.554	BB	306951	4551253	77.91%	4.927%
16	20.256	20.198	20.298	BB	254541	3790487	64.89%	4.103%
17	20.970	20.908	21.029	BB	189303	3329154	56.99%	3.604%
18	21.824	21.750	21.903	BB	146245	3406026	58.31%	3.687%
Sum of corrected areas:							92379906	

FG060122.M Mon Jun 27 09:41:26 2022