

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG062521\
 Data File : FG008429.D
 Signal(s) : FID1A.ch
 Acq On : 25 Jun 2021 23:53
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

mohammad
 6/28/2021 12:56:14 PM

Integration File: autoint1.e
 Quant Time: Jun 26 00:03:48 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG062421.M
 Quant Title :
 QLast Update : Thu Jun 24 09:09:15 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.769	6317093	48.283 ug/ml
Target Compounds			
1) N-OCTANE	2.361	7384363	52.796 ug/ml
2) N-DECANE	4.471	7606038	53.479 ug/ml
3) N-DODECANE	6.500	7714849	53.158 ug/ml
4) N-TETRADECANE	8.301	7721043	52.268 ug/ml
5) N-HEXADECANE	9.907	7579279	51.507 ug/ml
6) N-OCTADECANE	11.352	7649608	50.151 ug/ml
7) N-EICOSANE	12.666	7476963	49.198 ug/ml
8) N-DOCOSANE	13.867	7275714	48.438 ug/ml
10) N-TETRACOSANE	14.975	7104477	47.381 ug/ml
11) N-HEXACOSANE	15.996	6838377	46.505 ug/ml
12) N-OCTACOSANE	16.949	6544214	45.427 ug/ml
13) N-TRIACONTANE	17.837	6304651	44.778 ug/mlm
14) N-DOTRIACONTANE	18.671	6948914	51.923 ug/ml
15) N-TETRATRIACONTANE	19.457	5888623	47.439 ug/ml
16) N-HEXATRIACONTANE	20.197	5503758	49.816 ug/ml
17) N-OCTATRIACONTANE	20.953	4935286	51.262 ug/ml
18) N-TETRACONTANE	21.908	5297144	55.501 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG062521\
Data File : FG008429.D
Signal(s) : FID1A.ch
Acq On : 25 Jun 2021 23:53
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

mohammad
6/28/2021 12:56:14 PM

Integration File: autoint1.e
Quant Time: Jun 26 00:03:48 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG062421.M
Quant Title :
QLast Update : Thu Jun 24 09:09:15 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

