

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG062718\
 Data File : FG001498.D
 Signal(s) : FID1A.ch
 Acq On : 27 Jun 2018 22:14
 Operator : UA\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jun 28 01:39:44 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_G\Method\FG060118.M
 Quant Title :
 QLast Update : Fri Jun 01 15:28:31 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.536	5930569	48.512 ug/ml
Target Compounds			
2) N-DECANE	4.273	6783365	49.487 ug/ml
3) N-DODECANE	6.290	6877875	49.614 ug/ml
4) N-TETRADECANE	8.083	6924831	49.557 ug/ml
5) N-HEXADECANE	9.683	6962821	49.478 ug/ml
6) N-OCTADECANE	11.125	6975953	49.212 ug/ml
7) N-EICOSANE	12.435	6888648	48.919 ug/ml
8) N-DOCOSANE	13.635	6789525	48.446 ug/ml
10) N-TETRACOSANE	14.740	6692804	48.133 ug/ml
11) N-HEXACOSANE	15.762	5321956	38.418 ug/mlm
12) N-OCTACOSANE	16.712	6491305	47.046 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG062718\
Data File : FG001498.D
Signal(s) : FID1A.ch
Acq On : 27 Jun 2018 22:14
Operator : UA\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jun 28 01:39:44 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_G\Method\FG060118.M
Quant Title :
QLast Update : Fri Jun 01 15:28:31 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

