

rteres

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/29/2022
Supervised By :Ankita Jodhani 06/29/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG06282
Data File : FG010022.D
Signal(s) : FID1A.ch
Acq On : 28 Jun 2022 16:01
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG060122.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	1.836	1.761	1.909	BB	554265	5723495	82.47%	5.263%
2	4.343	4.309	4.390	BB	635948	6541843	94.26%	6.016%
3	6.522	6.488	6.566	BB	645353	6655316	95.90%	6.120%
4	8.352	8.313	8.398	BB	628295	6797847	97.95%	6.251%
5	9.961	9.919	10.008	BB	619376	6879847	99.13%	6.327%
6	11.405	11.358	11.438	BB	595679	6940167	100.00%	6.382%
7	12.718	12.667	12.768	BB	557583	6817611	98.23%	6.269%
8	13.918	13.883	13.965	VB	525972	6731225	96.99%	6.190%
9	14.820	14.763	14.846	BV	408422	5875113	84.65%	5.403%
10	15.023	14.947	15.076	BB	493575	6592791	94.99%	6.063%
11	16.046	15.988	16.097	BB	458058	6384828	92.00%	5.871%
12	16.998	16.943	17.051	BB	434384	6217678	89.59%	5.718%
13	17.886	17.828	17.940	BB	419574	6178571	89.03%	5.682%
14	18.720	18.660	18.767	BB	411331	5899928	85.01%	5.425%
15	19.505	19.443	19.556	BV	359196	5382329	77.55%	4.949%
16	20.246	20.185	20.288	BB	292340	4547739	65.53%	4.182%
17	20.961	20.907	21.025	BB	246944	4150333	59.80%	3.817%
18	21.813	21.727	21.905	BB	192288	4428320	63.81%	4.072%
Sum of corrected areas:						108744979		

FG060122.M Wed Jun 29 17:38:21 2022