

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG063022\
 Data File : FG010037.D
 Signal(s) : FID1A.ch
 Acq On : 30 Jun 2022 10:05
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 01 02:06:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG062922.M
 Quant Title :
 QLast Update : Thu Jun 30 04:15:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.815	6238482	52.465 ug/ml
Target Compounds			
1) N-OCTANE	1.831	6542802	52.625 ug/ml
2) N-DECANE	4.340	6707328	52.761 ug/ml
3) N-DODECANE	6.519	6808122	52.672 ug/ml
4) N-TETRADECANE	8.350	6940667	52.638 ug/ml
5) N-HEXADECANE	9.958	7006551	52.531 ug/ml
6) N-OCTADECANE	11.402	7112295	52.599 ug/ml
7) N-EICOSANE	12.714	7023041	52.429 ug/ml
8) N-DOCOSANE	13.914	7021785	52.480 ug/ml
10) N-TETRACOSANE	15.019	6967868	52.572 ug/ml
11) N-HEXACOSANE	16.041	6910473	52.704 ug/ml
12) N-OCTACOSANE	16.992	6770915	52.718 ug/ml
13) N-TRIACONTANE	17.883	6620690	52.618 ug/ml
14) N-DOTRIACONTANE	18.715	6216317	52.549 ug/ml
15) N-TETRATRIACONTANE	19.499	5608379	52.777 ug/ml
16) N-HEXATRIACONTANE	20.240	4694674	52.729 ug/ml
17) N-OCTATRIACONTANE	20.955	4122071	52.932 ug/ml
18) N-TETRACONTANE	21.802	4228777	53.966 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG063022\
Data File : FG010037.D
Signal(s) : FID1A.ch
Acq On : 30 Jun 2022 10:05
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 01 02:06:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG062922.M
Quant Title :
QLast Update : Thu Jun 30 04:15:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

