

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG063022\
 Data File : FG010041.D
 Signal(s) : FID1A.ch
 Acq On : 30 Jun 2022 12:06
 Operator : AJ\MA
 Sample : N3214-05
 Misc : 5 PPM SOIL LOQ
 ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 01 02:06:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG062922.M
 Quant Title :
 QLast Update : Thu Jun 30 04:15:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.813	2434200	20.472 ug/ml
Target Compounds			
1) N-OCTANE	1.831	566823	4.559 ug/ml
2) N-DECANE	4.338	527893	4.153 ug/ml
3) N-DODECANE	6.517	497744	3.851 ug/ml
4) N-TETRADECANE	8.347	492489	3.735 ug/ml
5) N-HEXADECANE	9.955	490830	3.680 ug/ml
6) N-OCTADECANE	11.399	488728	3.614 ug/ml
7) N-EICOSANE	12.709	500590	3.737 ug/ml
8) N-DOCOSANE	13.910	519851	3.885 ug/ml
10) N-TETRACOSANE	15.013	523125	3.947 ug/ml
11) N-HEXACOSANE	16.037	513797	3.919 ug/ml
12) N-OCTACOSANE	16.988	511842	3.985 ug/ml
13) N-TRIACONTANE	17.878	496828	3.949 ug/ml
14) N-DOTRIACONTANE	18.711	435378	3.680 ug/ml
15) N-TETRATRIACONTANE	19.495	333331	3.137 ug/ml
16) N-HEXATRIACONTANE	20.236	264006	2.965 ug/ml
17) N-OCTATRIACONTANE	20.952	225153	2.891 ug/ml
18) N-TETRACONTANE	21.800	205821	2.627 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG063022\
Data File : FG010041.D
Signal(s) : FID1A.ch
Acq On : 30 Jun 2022 12:06
Operator : AJ\MA
Sample : N3214-05
Misc : 5 PPM SOIL LOQ
ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 01 02:06:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG062922.M
Quant Title :
QLast Update : Thu Jun 30 04:15:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

