

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG063022\
 Data File : FG010046.D
 Signal(s) : FID1A.ch
 Acq On : 30 Jun 2022 16:49
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 01 02:07:41 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG062922.M
 Quant Title :
 QLast Update : Thu Jun 30 04:15:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.814	6020869	50.635 ug/ml
Target Compounds			
1) N-OCTANE	1.831	6329209	50.907 ug/ml
2) N-DECANE	4.339	6487608	51.033 ug/ml
3) N-DODECANE	6.519	6583317	50.933 ug/ml
4) N-TETRADECANE	8.350	6708456	50.877 ug/ml
5) N-HEXADECANE	9.957	6774137	50.788 ug/ml
6) N-OCTADECANE	11.402	6862774	50.754 ug/ml
7) N-EICOSANE	12.713	6782371	50.633 ug/ml
8) N-DOCOSANE	13.914	6771523	50.610 ug/ml
10) N-TETRACOSANE	15.019	6708764	50.617 ug/ml
11) N-HEXACOSANE	16.041	6642014	50.656 ug/ml
12) N-OCTACOSANE	16.993	6496376	50.580 ug/ml
13) N-TRIACONTANE	17.882	6353157	50.492 ug/ml
14) N-DOTRIACONTANE	18.714	5962835	50.406 ug/ml
15) N-TETRATRIACONTANE	19.498	5373193	50.564 ug/ml
16) N-HEXATRIACONTANE	20.240	4499660	50.538 ug/ml
17) N-OCTATRIACONTANE	20.953	3941622	50.615 ug/ml
18) N-TETRACONTANE	21.801	4010153	51.176 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG063022\
Data File : FG010046.D
Signal(s) : FID1A.ch
Acq On : 30 Jun 2022 16:49
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 01 02:07:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG062922.M
Quant Title :
QLast Update : Thu Jun 30 04:15:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

